



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/566,305	02/23/2006	James M. Davenport	SALTER P47AUSP1	1903
20210 7590 09/24/2009 DAVIS & BUJOLD, P.L.L.C. 112 PLEASANT STREET CONCORD, NH 03301				
EXAMINER				
OSTRUP, CLINTON T				
ART UNIT		PAPER NUMBER		
3771				
MAIL DATE		DELIVERY MODE		
09/24/2009		PAPER		

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

# Office Action Summary

**Application No.**

10/566,305

**Applicant(s)**

DAVENPORT ET AL.

**Examiner**

CLINTON OSTRUP

**Art Unit**

3771

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --  
**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 1 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☒ Responsive to communication(s) filed on 06 February 2009.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 1-43 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☐ Claim(s) \_\_\_\_\_ is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☒ Claim(s) 1-43 are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO/SE/US)  
Paper No(s)/Mail Date \_\_\_\_\_
- 4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date \_\_\_\_\_
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: \_\_\_\_\_

### **DETAILED ACTION**

1. Claims 1-43 are pending in this application.

#### ***Election/Restrictions***

2. Restriction is required under 35 U.S.C. 121 and 372.

This application contains the following inventions or groups of inventions which are not so linked as to form a single general inventive concept under PCT Rule 13.1.

3. In accordance with 37 CFR 1.499, applicant is required, in reply to this action, to elect a single invention to which the claims must be restricted.

Group I, claim(s) 1-19, drawn to a nasal cannula.

Group II, claim(s) 20-28, drawn to a respiratory therapy system.

Group III, claim(s) 29-35, drawn to a method of treating a patient with a sleep disorder.

Group IV, claim(s) 36-42, drawn to a diagnostic tool for measuring nasal cavity pressure.

Group V, claim(s) 43, drawn to a method of using a diagnostic tool.

4. The inventions listed as Groups I-V do not relate to a single general inventive concept under PCT Rule 13.1 because, under PCT Rule 13.2, they lack the same or corresponding special technical features for the following reasons:

5. A group of inventions is considered linked to form a single general inventive concept where there is a technical relationship among the inventions that involves at least one common or corresponding "special technical feature". The expression special technical features is defined as meaning those technical features that define the contribution which each claimed invention, considered as a whole, makes over the prior art.

6. In the instant case, the technical feature uniting groups I-V is the nasal cannula of group I. However, this nasal cannula is known, as demonstrated by Agdanowski et al (4,648,398).

7. Agdanowski discloses a nasal cannula (figure 1) for supplying a respiratory gas (12) to a patient (abstract), the nasal cannula (figure 1) comprising: a pair of supply lines (18 & 20) which each have a head (30 & 32) at one end thereof with a discharge opening (46 & 48) therein for discharging a respiratory gas (oxygen), and the opposite end (22) of each of the pair of supply lines being connectable (capable of being connected) to a respiratory gas source (12); wherein each head (30 & 32) is sized so that it would be snugly received and retained (col. 1, lines 47-51) within one of the nasal cavities of the patient (the hexagon shape which would inherently form voids at the concaved portion) which would form a sufficient leakage passage (via hexagon shapes and voids created therefrom), between a portion of inwardly facing nasal cavity skin of a patient and a portion of an exterior surface of the head (concaved portions of hexagon as shown in figure 4), to facilitate exhausting of any excess respiratory gas supplied to the patient (upon nasal exhalation) through the leakage passage and the voids created would also facilitate inhalation of any room air required in excess of the respiratory gas (oxygen) to be supplied to the patient.

8. Thus, unity of invention is lacking because there is no "special technical feature" uniting Groups I-V, since the technical feature uniting the claims does not make a contribution over the prior art.

9. Applicant is advised that the reply to this requirement to be complete must include (i) an election of a species or invention to be examined even though the requirement may be traversed (37 CFR 1.143) and (ii) identification of the claims encompassing the elected invention.

10. The election of an invention or species may be made with or without traverse. To preserve a right to petition, the election must be made with traverse. If the reply does not distinctly and specifically point out supposed errors in the restriction requirement, the election shall be treated as an election without traverse.

11. Applicant is reminded that upon the cancellation of claims to a non-elected invention, the inventorship must be amended in compliance with 37 CFR 1.48(b) if one or more of the currently named inventors is no longer an inventor of at least one claim remaining in the application. Any amendment of inventorship must be accompanied by a request under 37 CFR 1.48(b) and by the fee required under 37 CFR 1.17(i).

12. The examiner has required restriction between product and process claims. Where applicant elects claims directed to the product, and the product claims are subsequently found allowable, withdrawn process claims that depend from or otherwise require all the limitations of the allowable product claim will be considered for rejoinder. All claims directed to a nonelected process invention must require all the limitations of an allowable product claim for that process invention to be rejoined.

13. In the event of rejoinder, the requirement for restriction between the product claims and the rejoined process claims will be withdrawn, and the rejoined process claims will be fully examined for patentability in accordance with 37 CFR 1.104. Thus, to

be allowable, the rejoined claims must meet all criteria for patentability including the requirements of 35 U.S.C. 101, 102, 103 and 112. Until all claims to the elected product are found allowable, an otherwise proper restriction requirement between product claims and process claims may be maintained. Withdrawn process claims that are not commensurate in scope with an allowable product claim will not be rejoined. See MPEP § 821.04(b). Additionally, in order to retain the right to rejoinder in accordance with the above policy, applicant is advised that the process claims should be amended during prosecution to require the limitations of the product claims. **Failure to do so may result in a loss of the right to rejoinder.** Further, note that the prohibition against double patenting rejections of 35 U.S.C. 121 does not apply where the restriction requirement is withdrawn by the examiner before the patent issues. See MPEP § 804.01.

#### ***Conclusion***

14. Any inquiry concerning this communication or earlier communications from the examiner should be directed to CLINTON OSTRUP whose telephone number is (571)272-5559. The examiner can normally be reached on Monday-Friday.
15. If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Justine Yu can be reached on (571) 272-4835. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3771

16. Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Clinton Ostrup/  
Examiner, Art Unit 3771

/Justine R Yu/  
Supervisory Patent Examiner, Art Unit 3771